

REDUX DEVICE USE

OPEN



1

Open Redux by lifting tab on front of dryer.

2

Place up to two hearing aids in the Redux. Close the lid until it latches

3

Select Drying Mode: Rechargeable Battery, Non-Rechargeable Battery

4

Press start. Real-time drying data will begin to display at approximately 4 minutes until the process completes, then summary is displayed.

RE DUX